

## **Data Sheet**

# **O2S-FR-T5 Oxygen Sensors - Flange Mounted**

## **FEATURES**

- Oxygen pressure range 2 mbar 3bar
- Zirconium dioxide (ZrO<sub>2</sub>) sensing elements
- Non-consumption technology
- Integral heating element
- No need for temperature stabilisation
- No reference gas required
- High accuracy
- Linear output signal
- Operates with external interface boards



#### **SPECIFICATIONS**

Heater Voltage\*

Porous Cap (15mm)  $4.35V_{DC} \pm 0.1V_{DC}$  (1.85A) Stand by  $2V_{DC}$  (0.85A)

Pump resistance @  $700^{\circ}$ C\*\* <  $6k\Omega$ 

Permissible gas temperature -100 to 250°C

Gas flow rate 0 to 10 m/s

Repetitive permissible acceleration 5 g

Incidental permissible acceleration 30 g

## **ELECTRICAL CONNECTION**

- 1: Pump (Red)
- 2: Common (Black)
- 3: Heater (Yellow)
- 4: Sense (Blue)
- 5: Heater (Yellow)

Tel: +86-755-83289036

Fax: +86-755-83289052

E-mail: sales@isweek.com


<sup>\*</sup> It is important to measure the heater voltage as close to the sensor as possible due to voltage drops in the supply cable. Heater can also be operated with an equivalent a.c. or PWM signals.

 $<sup>^{\</sup>star\star}$  The constant current source used in the pump circuit should be designed to drive a load of up to  $6k\Omega$ 



# Data Sheet O2S-FR-T5 Oxygen Sensors - Flange Mounted

## **OUTLINE DRAWING**



Mass: 35g



## **Data Sheet**

## O2S-FR-T5 Oxygen Sensors - Flange Mounted

## PERFORMANCE CHARACTERISTIC

Characteristics	Min.	Тур.	Max.	Unit
Oxygen pressure range	2		3000	- mbar
Accuracy			5	
Internal operational temperature (full porous cap)		700 (4.35V <sub>DC</sub> )		°C
Response time (10-90% step, full porous cap)			4	s
Warm up time (prior to sensor operation)			100	
Warm up time (from stand by)			20	
Output stabilisation time		~180		

## **ORDERING INFORMATION**

Part number	Termination
O2S-FR-T5	0.4m Cable

## **WARNING**

**Personal Injury** 

DO NOT USE these products as safety or Emergengy Stop devices or in any other application Where failure of the product could result in Personal injury.

Failure to comply with these instructions could Result in death or serious injury.

## **CAUTION**

Do not exceed maximum ratings and ensure sensor is operated in accordance with all requirements of AN0043 Failure to comply with these instructions may result in product damage.

Tel: +86-755-83289036

Fax: +86-755-83289052

E-mail: sales@isweek.com